



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 9075			
SULFURIC ACID 0.6 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.6000 EQ/L	0.5995 - 0.6005	0.5997
Sample ID			
Tested 07/22/24			
Tester AFRE			
Recorded 07/22/24			
N.I.S.T. Traceable to SRM 723 THAM	YES YES		YES
Sample ID			
Tested 07/22/24			
Tester AFRE			
Recorded 07/22/24			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID			
Tested 07/22/24			
Tester AFRE			
Recorded 07/22/24			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 07/22/24			
Tester AFRE			
Recorded 07/22/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID			
Tested 07/22/24			
Tester AFRE			
Recorded 07/22/24			
DATE OF SAMPLING FOR TEST PURPOSES	YES MM/DD/YYYY		YES
Sample ID			
Tested 07/22/24			
Tester AFRE			
Recorded 07/22/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	20.0
Sample ID			
Tested 07/22/24			
Tester AFRE			
Recorded 07/22/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		07/22/2024
Sample ID			
Tested 07/22/24			
Tester AFRE			
Recorded 07/22/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	20.0
<b>DEG C</b>	<b>DEG C</b>		
Sample ID			
Tested	07/22/24		
Tester	AFRE		
Recorded	07/22/24		
<b>APPEARANCE AND COLOR</b>	<b>CLEAR WATER WHITE</b>		<b>PASS</b>
	<b>LIQUID</b>		
	<b>APPEARANCE</b>		
Sample ID			
Tested	07/22/24		
Tester	AFRE		
Recorded	07/22/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>INSTRUMENT</b>		<b>GC,VF6-17,M55,BID00488,PID M91</b>
Sample ID			
Tested	07/22/24		
Tester	AFRE		
Recorded	07/22/24		
<b>Actual Density</b>	<b>1.01762</b>	<b>1.01752 - 1.01772</b>	<b>1.01755</b>
	<b>GM/ML</b>		
Sample ID			
Tested	07/22/24		
Tester	AFRE		
Recorded	07/22/24		
<b>EXPIRATION DATE</b>	<b>MM/DD/YYYY</b>		<b>07/30/2026</b>
	<b>MM/DD/YYYY</b>		
Sample ID			
Tested	07/22/24		
Tester	AFRE		
Recorded	07/22/24		
<b>Lot #</b>	4071787		
<b>Made</b>	07/22/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**