

Page: 1

Date: 07/01/24 at 4:18 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

0 1 70 "

Customer PO #:

Packaged Product: 4353					
HYD	ROCHLORIC ACID 1.1 NORMAL				
Test	Target/UOM	Range	Result		
Actual Normality	1.1000	1.0995 - 1.1008	1.1005		
	EQ/L				
Sample ID					
Tested 07/01/24					
Tester AFRE					
Recorded 07/01/24					
QCSP5: STANDARDIZATION OF	YES		YES		
AQUEOUS ACIDS	METHOD				
Sample ID					
Tested 07/01/24					
Tester AFRE					
Recorded 07/01/24					
N.I.S.T. Traceable to SRM 84 (KH	P) Yes		YES		
	YES				
Sample ID					
Tested 07/01/24					
Tester AFRE					
Recorded 07/01/24					
PREPARED TO FORMULATION (YES		
FILE	YES				
Sample ID					
Tested 07/01/24					
Tester AFRE					
Recorded 07/01/24					
SAMPLE COLLECTION & HANDL			YES		
PER LAB040	YES				
Sample ID	.20				
Tested 07/01/24					
Tester AFRE	,				
Recorded 07/01/24 DATE OF SAMPLING FOR TEST	MM/DD/YYYY		07/01/2024		
			07/01/2024		
PURPOSES	MM/DD/YYYY				
Sample ID					
Tested 07/01/24					
Tester AFRE					
Recorded 07/01/24		40.0	22.2		
AMBIENT TEMPERATURE DURIN		10.0 - 35.0	20.0		
SAMPLING DEGC	DEG C				
Sample ID					
Tested 07/01/24					
Tester AFRE					
Recorded 07/01/24					
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		07/01/2024		
Sample ID					
Tested 07/01/24					
Tester AFRE					
Recorded 07/01/24					
110001ueu 07/01/24	•				

07/01/24

Made

Page: 2

Date: 07/01/24 at 4:18 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

AMBIENT TEST TEMPER	RATURE	20.0	15.0 - 30.0	20.0
DEG C		DEG C		
Sample ID				
Tested	07/01/24			
Tester	AFRE			
Recorded	07/01/24			
APPEARANCE AND COL	.OR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	07/01/24			
Tester	AFRE			
Recorded	07/01/24			
INSTRUMENTS USED DI	JRING	INSTRUMENT		M103,VF5-L9
PREPARATION				
Sample ID				
Tested	07/01/24			
Tester	AFRE			
Recorded	07/01/24			
EXPIRATION DATE		MM/DD/YYYY		07/30/2026
		MM/DD/YYYY		
Sample ID				
Tested	07/01/24			
Tester	AFRE			
Recorded	07/01/24			
Actual Density		1.01764	1.01760 - 1.01769	1.01766
		GM/ML		
Sample ID				
Tested	07/01/24			
Tester	AFRE			
Recorded	07/01/24			
Lot # 406270	06			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1000000

Quality Manager