



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: OXY010F			
LEVEL 5 CELL LIQUOR STANDARD			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
CALCULATED GPL SODIUM CHLORIDE	200.00 GPL	199.98 - 200.02	200.00
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
CALCULATED GPL SODIUM HYDROXIDE	100.0 GPL	99.9 - 100.1	100.0
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
CALCULATED GPL SODIUM CHLORATE	1.0000 GPL	0.9995 - 1.0005	1.0000
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,M108
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
Actual Density	1.21661 GM/ML	1.21526 - 1.21795	1.21655
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		07/30/25
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

SHELF LIFE
TIME

1 YEAR

Sample ID
 Tested 07/31/24
 Tester JTRA
 Recorded 07/31/24

Lot # 4072667
Made 07/31/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager