



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: OXY008F			
LEVEL 3 CELL LIQUOR STANDARD			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
CALCULATED GPL SODIUM CHLORIDE	150.00 GPL	149.98 - 150.02	150.00
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
CALCULATED GPL SODIUM HYDROXIDE	160.0 GPL	159.9 - 160.1	160.0
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
CALCULATED GPL SODIUM CHLORATE	0.5000 GPL	0.4995 - 0.5004	0.5000
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,M108
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
Actual Density	1.23555 GM/ML	1.22827 - 1.24284	1.23549
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		07/30/25
Sample ID			
Tested 07/31/24			
Tester JTRA			
Recorded 07/31/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested	07/31/24	
Tester	JTRA	
Recorded	07/31/24	
Lot #	4072666	
Made	07/31/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager