



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9040			
SULFURIC ACID 0.1 NORMAL NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		07/30/2024
	MM/DD/YYYY		
Sample ID			
Tested 07/30/24			
Tester ENOR			
Recorded 07/30/24			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	19.1
DEG C	DEG C		
Sample ID			
Tested 07/30/24			
Tester ENOR			
Recorded 07/30/24			
Actual Normality	0.1000	0.0995 - 0.1005	0.1001
	EQ/L		
Sample ID			
Tested 07/30/24			
Tester ENOR			
Recorded 07/30/24			
QCSP5: STANDARDIZATION OF	YES		YES
AQUEOUS ACIDS	METHOD		
Sample ID			
Tested 07/30/24			
Tester ENOR			
Recorded 07/30/24			
N.I.S.T. Traceable to SRM 84	Yes		YES
	YES		
Sample ID			
Tested 07/30/24			
Tester ENOR			
Recorded 07/30/24			
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/30/24			
Tester ENOR			
Recorded 07/30/24			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 07/30/24			
Tester ENOR			
Recorded 07/30/24			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		07/30/2024
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 07/30/24			
Tester ENOR			
Recorded 07/30/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	20.0
Sample ID			
Tested	07/30/24		
Tester	ENOR		
Recorded	07/30/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	07/30/24		
Tester	ENOR		
Recorded	07/30/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		GC,T-57,M55,BID08446,PIDM9 1
Sample ID			
Tested	07/30/24		
Tester	ENOR		
Recorded	07/30/24		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		2 YEARS
Sample ID			
Tested	07/30/24		
Tester	ENOR		
Recorded	07/30/24		
Lot #	4072475		
Made	07/30/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager