



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: LY386F			
SULFUR STANDARD 10.0 PPM W/W IN ISOCTANE			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/01/24			
Tester JTRA			
Recorded 08/01/24			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 08/01/24			
Tester JTRA			
Recorded 08/01/24			
Calculated ppm Sulfur	10.00 PPM	9.80 - 10.20	10.00
Sample ID			
Tested 08/01/24			
Tester TAUS			
Recorded 08/01/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P176,VF1-Z24
Sample ID			
Tested 08/01/24			
Tester JTRA			
Recorded 08/01/24			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		08/30/2025
Sample ID			
Tested 08/01/24			
Tester JTRA			
Recorded 08/01/24			
Lot #	4072948		
Made	08/01/24		

Quality Manager