Page: 1

Date: 07/31/24 at 2:40 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	CG048F		
	Total nitrogen standard, 10	.0 ppm ASTM D8083	
Test	Target/UOM	Range	Result
PREPARED TO FORMULA	TION ON YES		YES
FILE	YES		
Sample ID			
Tested 0	7/31/24		
Tester A	ROS		
Recorded 0	7/31/24		
Calculated ppm (ug/g) Nitr	rogen 10.00	9.95 - 10.05	10.00
	UG/G (PPM)		
Sample ID			
Tested 0	7/31/24		
Tester A	ROS		
Recorded 0	7/31/24		
APPEARANCE AND COLO	R CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 0	7/31/24		
Tester A	ROS		
Recorded 0	7/31/24		
INSTRUMENTS USED DUF	RING INSTRUMENT		P63,VF3-J15
PREPARATION			
Sample ID			
Tested 0	7/31/24		
Tester A	ROS		
Recorded 0	7/31/24		
EXPIRATION DATE	MM/DD/YYYY		11/30/2024
	MM/DD/YYYY		
Sample ID			
•	7/31/24		
Tester A	ROS		
Recorded 0	7/31/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I