



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1490			
BUFFER REFERENCE pH 6.86			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 08/07/24			
Tester BPHI			
Recorded 08/07/24			
Actual pH of product	6.860	6.850 - 6.870	6.862
	PH		
Sample ID			
Tested 08/07/24			
Tester BPHI			
Recorded 08/07/24			
QCSP 20 PH CALIBRATION METHODS	YES		YES
	METHOD		
Sample ID			
Tested 08/07/24			
Tester BPHI			
Recorded 08/07/24			
Actual Temp	25.0	24.5 - 25.5	25.0
	DEG C		
Sample ID			
Tested 08/07/24			
Tester BPHI			
Recorded 08/07/24			
NIST TRACEABLE TO SRM186I AND 186II	YES		YES
	YES		
Sample ID			
Tested 08/07/24			
Tester BPHI			
Recorded 08/07/24			
PH NIST TRACEABLE TO SRM 185	YES		YES
	YES		
Sample ID			
Tested 08/07/24			
Tester BPHI			
Recorded 08/07/24			
PH NIST TRACEABLE TO SRM 187	YES		YES
	YES		
Sample ID			
Tested 08/07/24			
Tester BPHI			
Recorded 08/07/24			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 08/07/24			
Tester BPHI			
Recorded 08/07/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M103,T-36,M117
Sample ID		
Tested 08/07/24		
Tester BPHI		
Recorded 08/07/24		
EXPIRATION DATE	MM/DD/YY	08/30/25
	MM/DD/YYYY	
Sample ID		
Tested 08/07/24		
Tester BPHI		
Recorded 08/07/24		
SHELF LIFE OF PRODUCT	SHELF LIFE	1 YEAR
	TIME	
Sample ID		
Tested 08/07/24		
Tester BPHI		
Recorded 08/07/24		
Lot # 4072970		
Made 08/07/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager