

Page: 1

Date: 08/07/24 at 5:05 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Customer PO #:					
Packaged Product: 9079					
SULFURIO	ACID 0.5 NORMAL NIST TRACEABLE (AUTO-TITRATOR)				
Test	Target/UOM	Range	Result		
PREPARED TO FORMULATION ON	YES	-	YES		
FILE	YES				
Sample ID					
Tested 08/07/24					
Tester SHAR					
Recorded 08/07/24					
SAMPLE COLLECTION & HANDLING	YES		YES		
PER LAB040	YES				
Sample ID					
Tested 08/07/24					
Tester SHAR					
Recorded 08/07/24					
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		08/07/2024		
PURPOSES	MM/DD/YYYY				
Sample ID					
Tested 08/07/24					
Tester SHAR					
Recorded 08/07/24					
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	21.0		
SAMPLING DEGC	DEG C				
Sample ID					
Tested 08/07/24					
Tester SHAR					
Recorded 08/07/24					
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/07/2024		
	MM/DD/YYYY				
Sample ID					
Tested 08/07/24					
Tester SHAR					
Recorded 08/07/24					
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	21.0		
DEG C	DEG C		-		
Sample ID	 -				
Tested 08/07/24					
Tester SHAR					
Recorded 08/07/24					
NORMALITY @ 20 DEG C	0.5000	0.4995 - 0.5005	0.5004		
	EQUIV/LITER	2222 0.0000	2.000 !		
Sample ID					
Tested 08/07/24					
Tester SHAR					
Recorded 08/07/24					
QCSP58:STANDARDIZATION ACID	YES		YES		
SOLN TO POTE	QCSP58		ILO		
	QUOF JO				
Sample ID					
Tested 08/07/24					
Tester SHAR					
Recorded 08/07/24					

08/07/24

Made

Page: 2

Date: 08/07/24 at 5:05 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	08/07/24			
Tester	SHAR			
Recorded	08/07/24			
INSTRUMENTS USED D	URING	INSTRUMENT		GC,EF6-AC3,M55,BURET:079,
PREPARATION				M91
Sample ID				
Tested	08/07/24			
Tester	SHAR			
Recorded	08/07/24			
Actual Density		1.01445	1.01433 - 1.01457	1.01457
-		GM/ML		
Sample ID				
Tested	08/07/24			
Tester	SHAR			
Recorded	08/07/24			
EXPIRATION DATE		MM/DD/YY		08/30/26
		MM/DD/YYYY		
Sample ID				
Tested	08/07/24			
Tester	SHAR			
Recorded	08/07/24			
N.I.S.T. Traceable to SRI		YES		YES
		YES		-
Sample ID				
Tested	08/07/24			
Tester	SHAR			
Recorded	08/07/24			
Lot # 40807				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Mullity Manager