



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7350			
POTASSIUM IODIDE-IODATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.1000	0.0995 - 0.1005	0.0997
	EQ/L		
Sample ID			
Tested	08/06/24		
Tester	SHAR		
Recorded	08/06/24		
Actual Density	1.02693	1.02581 - 1.02805	1.02659
	GM/ML		
Sample ID			
Tested	08/06/24		
Tester	SHAR		
Recorded	08/06/24		
QCSP33:POTASSIUM IODATE-IODIDE TITRANT	YES		YES
	QCSP33		
Sample ID			
Tested	08/06/24		
Tester	SHAR		
Recorded	08/06/24		
N.I.S.T. Traceable to SRM 136	Yes		YES
	YES		
Sample ID			
Tested	08/06/24		
Tester	SHAR		
Recorded	08/06/24		
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested	08/06/24		
Tester	SHAR		
Recorded	08/06/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	08/06/24		
Tester	SHAR		
Recorded	08/06/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,T-192,M55,BID079,PIDP 90
Sample ID			
Tested	08/06/24		
Tester	SHAR		
Recorded	08/06/24		
SHELF LIFE OF PRODUCT	1 YEAR TIME		1 YEAR
Sample ID			
Tested	08/06/24		
Tester	SHAR		
Recorded	08/06/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 08/06/24 at 3:19 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE

MM/DD/YYYY

08/30/2025

MM/DD/YYYY

Sample ID

Tested 08/06/24

Tester AFRE

Recorded 08/06/24

Lot # 4072602

Made 08/06/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager