



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 9090			
SULFURIC ACID 1.0 NORMAL NIST TRACEABLE			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 08/05/24			
Tester JENC			
Recorded 08/05/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID	YES		
Tested 08/05/24			
Tester JENC			
Recorded 08/05/24			
<b>Actual Normality</b>	1.0000 EQ/L	0.9995 - 1.0005	1.0003
Sample ID			
Tested 08/05/24			
Tester JENC			
Recorded 08/05/24			
<b>QCSP5: STANDARDIZATION OF AQUEOUS ACIDS</b>	YES METHOD		YES
Sample ID			
Tested 08/05/24			
Tester JENC			
Recorded 08/05/24			
<b>N.I.S.T. Traceable to SRM 723 THAM</b>	YES		YES
Sample ID	YES		
Tested 08/05/24			
Tester JENC			
Recorded 08/05/24			
<b>Actual Density</b>	1.03000 GM/ML	1.02978 - 1.03021	1.03008
Sample ID			
Tested 08/05/24			
Tester JENC			
Recorded 08/05/24			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/05/24			
Tester JENC			
Recorded 08/05/24			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		T-532,GC,T-71,M55,BID02259, PIDM91
Sample ID			
Tested 08/05/24			
Tester JENC			
Recorded 08/05/24			




**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	08/30/26
Sample ID		
Tested	08/05/24	
Tester	JENC	
Recorded	08/05/24	
<b>SHELF LIFE OF PRODUCT</b>	<b>SHELF LIFE</b>	<b>2 YEARS</b>
	<b>TIME</b>	
Sample ID		
Tested	08/05/24	
Tester	JENC	
Recorded	08/05/24	
<b>Lot #</b>	40729102	
<b>Made</b>	08/05/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



**Quality Manager**