



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3650			
EDTA TITRANT 0.0575 MOLAR (0.115N)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 08/07/24			
Tester ENOR			
Recorded 08/07/24			
Actual Molarity	0.0575 MOL/L	0.0570 - 0.0580	0.0576
Sample ID			
Tested 08/07/24			
Tester ENOR			
Recorded 08/07/24			
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES QCSP4		YES
Sample ID			
Tested 08/07/24			
Tester ENOR			
Recorded 08/07/24			
N.I.S.T. Traceable to SRM 915	YES YES		YES
Sample ID			
Tested 08/07/24			
Tester ENOR			
Recorded 08/07/24			
Actual Density	1.00888 GM/ML	1.00837 - 1.00939	1.00882
Sample ID			
Tested 08/07/24			
Tester ENOR			
Recorded 08/07/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/07/24			
Tester ENOR			
Recorded 08/07/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,T-38,M55,BID06969,PID M91
Sample ID			
Tested 08/07/24			
Tester ENOR			
Recorded 08/07/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		08/30/26
Sample ID			
Tested 08/07/24			
Tester ENOR			
Recorded 08/07/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 08/07/24 at 7:09 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

2 YEARS
TIME

2 YEARS

Sample ID

Tested 08/07/24

Tester ENOR

Recorded 08/07/24

Lot # 4072942

Made 08/07/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager