



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: B0810			
BATH FLUID SILICONE HI-TEMP			
Test	Target/UOM	Range	Result
VISCOSITY AT 25 DEG C	90-110 CST		101.7
Sample ID			
Tested	08/08/24		
Tester	TBUR		
Recorded	08/08/24		
APPEARANCE	AMBER, TRANSPARENT FLUID		PASS
Sample ID			
Tested	08/08/24		
Tester	TBUR		
Recorded	08/08/24		
SPECIFIC GRAVITY	0.975 SPEC GRAV	0.972 - 0.977	0.966
Sample ID			
Tested	08/08/24		
Tester	TBUR		
Recorded	08/08/24		
FLASH POINT, CC DEG F	<= 610 DEG F DEG F		PASS
Sample ID			
Tested	08/08/24		
Tester	TBUR		
Recorded	08/08/24		
VOLATILE CONTENT 24 HR AT 150 DEG C	0.0-<2.0 %		<0.50
Sample ID			
Tested	08/08/24		
Tester	TBUR		
Recorded	08/08/24		
MANUFACTURE DATE	MM/DD/YY MM/DD/YYYY		05/25/2023
Sample ID			
Tested	08/08/24		
Tester	TBUR		
Recorded	08/08/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		05/25/2026
Sample ID			
Tested	08/08/24		
Tester	TBUR		
Recorded	08/08/24		
SHELF LIFE OF PRODUCT	3 YEARS TIME		3 YEARS
Sample ID			
Tested	08/08/24		
Tester	TBUR		
Recorded	08/08/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 08/08/24 at 9:34 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	422005
Made	08/07/24

Quality Manager