



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8975			
SULFIDE STANDARD 1000 PPM W/W (1ML = 1MG S2-)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 08/05/24			
Tester SHAR			
Recorded 08/05/24			
PPM SULFIDE TO 1 DECIMAL PLACE	1000.0	990.0 - 1010.0	991.3
Sample ID	PPM		
Tested 08/05/24			
Tester SHAR			
Recorded 08/05/24			
QCSP56:STANDARDIZATION SODIUM SULFIDE SO	YES		YES
Sample ID	QCSP56		
Tested 08/05/24			
Tester SHAR			
Recorded 08/05/24			
Actual Density	1.00608	1.00497 - 1.00719	1.00702
Sample ID	GM/ML		
Tested 08/05/24			
Tester SHAR			
Recorded 08/05/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/05/24			
Tester SHAR			
Recorded 08/05/24			
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT		M103,VF12-98,M55,BID079,PI DP108
Sample ID			
Tested 08/05/24			
Tester SHAR			
Recorded 08/05/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		10/10/24
Sample ID			
Tested 08/05/24			
Tester SHAR			
Recorded 08/05/24			
SHELF LIFE OF PRODUCT	2 MONTHS TIME		2 MONTHS
Sample ID			
Tested 08/05/24			
Tester SHAR			
Recorded 08/05/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

PREPARED TO STANDARD	YES	YES
METHODS 4500-S2		
Sample ID		
Tested	08/05/24	
Tester	SHAR	
Recorded	08/05/24	
Lot #	4080501	
Made	08/05/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager