



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8550			
SILVER AA STANDARD 1000 PPM (NIST)			
Test	Target/UOM	Range	Result
CONCENTRATION, MG/L	1000 PPM	990 - 1010	1000
Sample ID			
Tested 08/23/24			
Tester TBUR			
Recorded 08/23/24			
NIST TRACEABLE TO SRM 999	YES YES		YES
Sample ID			
Tested 08/23/24			
Tester TBUR			
Recorded 08/23/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		02/09/2029
Sample ID			
Tested 08/23/24			
Tester TBUR			
Recorded 08/23/24			
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		5 YEARS
Sample ID			
Tested 08/23/24			
Tester TBUR			
Recorded 08/23/24			
MANUFACTURE DATE	MM/DD/YY MM/DD/YYYY		02/09/2024
Sample ID			
Tested 08/23/24			
Tester TBUR			
Recorded 08/23/24			
Lot #	423402		
Made	08/21/24		

Richard [Signature]
Quality Manager