



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> ERL125F			
	SILVER NITRATE 0.1 NORMAL IN ISOPROPANOL		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 08/29/24			
Tester JENC			
Recorded 08/29/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 08/29/24			
Tester JENC			
Recorded 08/29/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		08/29/2024
Sample ID			
Tested 08/29/24			
Tester JENC			
Recorded 08/29/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	19.8
Sample ID			
Tested 08/29/24			
Tester JENC			
Recorded 08/29/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		08/29/2024
Sample ID			
Tested 08/29/24			
Tester JENC			
Recorded 08/29/24			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	19.8
Sample ID			
Tested 08/29/24			
Tester JENC			
Recorded 08/29/24			
<b>Actual Normality</b>	0.1000 EQ/L	0.0995 - 0.1005	0.1011
Sample ID			
Tested 08/29/24			
Tester JENC			
Recorded 08/29/24			
<b>QCSP53: SILVER NITRATE STD ON AUTO-TITRA</b>	YES METHOD		YES
Sample ID			
Tested 08/29/24			
Tester JENC			
Recorded 08/29/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>N.I.S.T. Traceable to SRM 919 NaCl</b>	YES		YES
	YES		
Sample ID			
Tested	08/29/24		
Tester	JENC		
Recorded	08/29/24		
<b>PREPARED TO ASTM D3227</b>	YES		YES
	YES		
Sample ID			
Tested	08/29/24		
Tester	JENC		
Recorded	08/29/24		
<b>Actual Density</b>	0.82561	0.82176 - 0.82947	0.82639
	GM/ML		
Sample ID			
Tested	08/29/24		
Tester	JENC		
Recorded	08/29/24		
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested	08/29/24		
Tester	JENC		
Recorded	08/29/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		CUBE,M55,M52,P18
Sample ID			
Tested	08/29/24		
Tester	JENC		
Recorded	08/29/24		
<b>LOT NUMBER OF ORIGINAL MATERIAL</b>	LOT #		4062808
Sample ID			
Tested	08/29/24		
Tester	JENC		
Recorded	08/29/24		
<b>EXPIRATION DATE</b>	MM/DD/YY		09/10/24
	MM/DD/YYYY		
Sample ID			
Tested	08/29/24		
Tester	JENC		
Recorded	08/29/24		
<b>SHELF LIFE OF PRODUCT</b>	1 WEEK		1 WEEK
	TIME		
Sample ID			
Tested	08/29/24		
Tester	JENC		
Recorded	08/29/24		

**Lot #** 4082608  
**Made** 08/29/24



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 08/29/24 at 2:19 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**