

Made 08/26/24

Page: 1

Date: 08/26/24 at 8:33 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

			110#:	
Packaged Product:	6925			
	CONDUCT	CIVITY STANDARD 2,500 uMHC)S	
Test		Target/UOM	Range	Result
PREPARED TO FORMULAT	TION ON	YES		YES
FILE		YES		
Sample ID				
	3/26/24			
	PHI			
	3/26/24			
CONDUCTIVITY @25 C.		2500	2499 - 2501	2500
		US/CM		
Sample ID				
Tested 08	3/26/24			
	PHI			
Recorded 08	3/26/24			
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested 08	3/26/24			
Tester BF	PHI			
Recorded 08	3/26/24			
TRACEABLE TO N.I.S.T. SF	RM 3191	YES		YES
		YES		
Sample ID				
Tested 08	3/26/24			
Tester BF	PHI			
Recorded 08	3/26/24			
APPEARANCE AND COLOR	3	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
	3/26/24			
	PHI			
	3/26/24			
INSTRUMENTS USED DUR		INSTRUMENT		M92,T-229,BOTTLE
PREPARATION				•
Sample ID				
	3/26/24			
	PHI			
	3/26/24			
EXPIRATION DATE	== '	MM/DD/YY		08/30/25
		MM/DD/YYYY		
Sample ID				
·	3/26/24			
	720/24 PHI			
	3/26/24			
Lot # 4082246				
LOC# 1002210				



Page: 2

Date: 08/26/24 at 8:33 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager