

Page: 1

Date: 08/27/24 at 7:09 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 69	81		
CO	NDUCTIVITY STANDARD 18,000	UMHOS	
Test	Target/UOM	Range	Result
TRACEABLE TO N.I.S.T. SRM 3	191 YES	<u> </u>	YES
	YES		
Sample ID			
Tested 08/27/2	24		
Tester BPHI			
Recorded 08/27/2			
PREPARED TO FORMULATION			YES
FILE	YES		
Sample ID			
Tested 08/27/	24		
Tester BPHI			
Recorded 08/27/			
CONDUCTIVITY @25 C.	18000	17900 - 18100	18000
	US/CM		
Sample ID			
Tested 08/27/	24		
Tester BPHI			
Recorded 08/27/	24		
QCSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 08/27/	24		
Tester BPHI			
Recorded 08/27/	24		
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 08/27/	24		
Tester BPHI			
Recorded 08/27/	24		
INSTRUMENTS USED DURING	INSTRUMENT		M92,T-234
PREPARATION			
Sample ID			
Tested 08/27/	24		
Tester BPHI			
Recorded 08/27/	24		
EXPIRATION DATE	MM/DD/YYYY		08/30/2025
	MM/DD/YYYY		33/33/2323
Sample ID			
Tested 08/27/	24		
Tester BPHI	_ _		
Recorded 08/27/	24		
Made 08/27/24			



Page: 2

Date: 08/27/24 at 7:09 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager