

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	THE521F			
	Phosphor	Reagent		
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	08/28/24			
Tester	BPHI			
Recorded	08/28/24			
Calculated v/v% Phosphoric Acid		15.79	15.74 - 15.84	15.79
		V/V%		
Sample ID				
	08/28/24			
	BPHI			
Recorded	08/28/24			
Calculated v/v% Sulfuric Acid		10.53	10.48 - 10.58	10.53
		V/V%		
Sample ID				
	08/28/24			
	BPHI			
	08/28/24			
APPEARANCE AND COL	OR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
	08/28/24			
	BPHI			
	08/28/24			
INSTRUMENTS USED DU	JRING	INSTRUMENT		T-228,GC,VF5-L5,VF4-N22,VF
PREPARATION				3-J10,T-
Sample ID)
•	08/28/24			
	BPHI			
	08/28/24			
EXPIRATION DATE		MM/DD/YYYY		08/30/2026
		MM/DD/YYYY		
Sample ID				
	08/29/24			
	BPHI			
	08/29/24			
Lot # 408223				
Made 08/28/				
	27			



Page: 2 Date: 08/29/24 at 2:56 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager