



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 08/29/24 at 11:14 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ASI009F

BROMINE INDEX STANDARD 10 MG/100G IN TOLUENE

Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 08/29/24			
Tester AROS			
Recorded 08/29/24			
BROMINE NUMBER MG BR2/100 GM	10.00	9.50 - 10.50	10.00
	MG/100G		
Sample ID			
Tested 08/29/24			
Tester AROS			
Recorded 08/29/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested 08/29/24			
Tester AROS			
Recorded 08/29/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112
Sample ID			
Tested 08/29/24			
Tester AROS			
Recorded 08/29/24			
EXPIRATION DATE	MM/DD/YYYY		02/28/2025
	MM/DD/YYYY		
Sample ID			
Tested 08/29/24			
Tester AROS			
Recorded 08/29/24			
Lot #	4082959		
Made	08/29/24		

Quality Manager