



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4750-1L
 IODINE 0.1 NORMAL SOLN. CERTIFIED (1LG)

Lot # 4083104

Chemical Product: 4750
 IODINE 0.1 NORMAL SOLN. CERTIFIED

Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 08/15/24			
Tester SHAR			
Recorded 08/15/24			
Actual Normality	0.1000	0.0995 - 0.1005	0.1003
	EQ/L		
Sample ID			
Tested 08/15/24			
Tester SHAR			
Recorded 08/15/24			
APPEARANCE AND COLOR	DARK BROWN LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/15/24			
Tester SHAR			
Recorded 08/15/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,T142
Sample ID			
Tested 08/15/24			
Tester SHAR			
Recorded 08/15/24			
EXPIRATION DATE	MM/DD/YY		08/30/25
	MM/DD/YYYY		
Sample ID			
Tested 08/15/24			
Tester SHAR			
Recorded 08/15/24			
SHELF LIFE OF PRODUCT	1 YEAR		1 YEAR
	TIME		
Sample ID			
Tested 08/15/24			
Tester SHAR			
Recorded 08/15/24			
Actual Density	1.03669	1.03439 - 1.03900	1.03700
	GM/ML		
Sample ID			
Tested 08/15/24			
Tester SHAR			
Recorded 08/15/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

QCSP12:STANDARDIZATION OF
 IODINE SOLUTIO

YES
 QCSP12

YES

Sample ID
 Tested 08/15/24
 Tester SHAR
 Recorded 08/15/24

Made 08/15/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager