



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: DC662F			
	HYDROCHLORIC ACID 0.01 NORMAL (MR/THAM)		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 08/30/24			
Tester ENOR			
Recorded 08/30/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 08/30/24			
Tester ENOR			
Recorded 08/30/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		08/30/2024
Sample ID	MM/DD/YYYY		
Tested 08/30/24			
Tester ENOR			
Recorded 08/30/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	18.3
Sample ID	DEG C		
Tested 08/30/24			
Tester ENOR			
Recorded 08/30/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/30/2024
Sample ID	MM/DD/YYYY		
Tested 08/30/24			
Tester ENOR			
Recorded 08/30/24			
Actual Normality	0.0100	0.0095 - 0.0105	0.0095
Sample ID	EQ/L		
Tested 08/30/24			
Tester ENOR			
Recorded 08/30/24			
N.I.S.T. Traceable to SRM 723 THAM	YES		YES
Sample ID	YES		
Tested 08/30/24			
Tester ENOR			
Recorded 08/30/24			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
Sample ID	APPEARANCE		
Tested 08/30/24			
Tester ENOR			
Recorded 08/30/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	GC,T-22,M55,BURET:06969,M 91
Sample ID		
Tested 08/30/24		
Tester ENOR		
Recorded 08/30/24		
Actual Density	0.99842 GM/ML	0.99833 - 0.99851 0.99841
Sample ID		
Tested 08/30/24		
Tester ENOR		
Recorded 08/30/24		
AMBIENT TEST TEMPERATURE DEG C	20.0 DEG C	15.0 - 30.0 20.0
Sample ID		
Tested 08/30/24		
Tester ENOR		
Recorded 08/30/24		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	08/30/2025
Sample ID		
Tested 08/30/24		
Tester ENOR		
Recorded 08/30/24		
Lot # 4082710		
Made 08/30/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager