

Page: 1

Date: 09/05/24 at 9:11 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product				
	COLOR	STANDARD 5 UNITS (APHA)		
Test		Target/UOM	Range	Result
Prepared to formulatio	n on File	Yes		YES
		YES		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
Actual Color, APHA		5	4 - 6	5
		COLOR		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
APPEARANCE AND CO		CLEAR PALE YELLOW		PASS
		APPEARANCE		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
INSTRUMENTS USED I		INSTRUMENT		P173,P69,VF12-98,M44
PREPARATION	Joining .	ii to i i tome. Ti		1 170,1 00,11 12 00,111 1
Sample ID				
Tested	09/05/24			
Tester	09/05/24 AROS			
Recorded	09/05/24			
EXPIRATION DATE	09/05/24	MM/DD/YY		09/30/25
EXPIRATION DATE				09/30/23
0 1 15		MM/DD/YYYY		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24	4 MEAD		1.7/54.5
SHELF LIFE OF PRODUCT		1 YEAR		1 YEAR
		TIME		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
Lot # 4090	429			
Made 09/0				
IVIAUE 0370	-,			



Page: 2

Date: 09/05/24 at 9:11 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager