



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: 9113 | | | |
|--|-------------|---------------|------------|
| SULFURIC ACID 6 NORMAL WITH MNSO4 40 MG/L | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | YES | | |
| Tested 09/03/24 | | | |
| Tester ENOR | | | |
| Recorded 09/03/24 | | | |
| SAMPLE COLLECTION & HANDLING PER LAB040 | YES | | YES |
| Sample ID | YES | | |
| Tested 09/03/24 | | | |
| Tester ENOR | | | |
| Recorded 09/03/24 | | | |
| DATE OF SAMPLING FOR TEST PURPOSES | MM/DD/YYYY | | 09/03/2024 |
| Sample ID | MM/DD/YYYY | | |
| Tested 09/03/24 | | | |
| Tester ENOR | | | |
| Recorded 09/03/24 | | | |
| AMBIENT TEMPERATURE DURING SAMPLING DEGC | 25.0 | 10.0 - 35.0 | 17.6 |
| Sample ID | DEG C | | |
| Tested 09/03/24 | | | |
| Tester ENOR | | | |
| Recorded 09/03/24 | | | |
| DATE OF TESTING OF SAMPLE | MM/DD/YYYY | | 09/03/2024 |
| Sample ID | MM/DD/YYYY | | |
| Tested 09/03/24 | | | |
| Tester ENOR | | | |
| Recorded 09/03/24 | | | |
| AMBIENT TEST TEMPERATURE DEG C | 20.0 | 15.0 - 30.0 | 20.0 |
| Sample ID | DEG C | | |
| Tested 09/03/24 | | | |
| Tester ENOR | | | |
| Recorded 09/03/24 | | | |
| NORMALITY | 6.000 | 5.970 - 6.030 | 5.987 |
| Sample ID | EQUIV/LITER | | |
| Tested 09/03/24 | | | |
| Tester ENOR | | | |
| Recorded 09/03/24 | | | |
| QCSP5: STANDARDIZATION OF AQUEOUS ACIDS | YES | | YES |
| Sample ID | METHOD | | |
| Tested 09/03/24 | | | |
| Tester ENOR | | | |
| Recorded 09/03/24 | | | |



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 723 THAM

YES

YES

YES

Sample ID

Tested 09/03/24

Tester ENOR

Recorded 09/03/24

SUITABLE FOR TOC USE

PASS

PASS

PASS

Sample ID

Tested 09/03/24

Tester ENOR

Recorded 09/03/24

APPEARANCE AND COLOR

CLEAR WATER WHITE

PASS

APPEARANCE

Sample ID

Tested 09/03/24

Tester ENOR

Recorded 09/03/24

INSTRUMENTS USED DURING

INSTRUMENT

T161,T172

PREPARATION

Sample ID

Tested 09/03/24

Tester ENOR

Recorded 09/03/24

EXPIRATION DATE

MM/DD/YYYY

03/30/2026

MM/DD/YYYY

Sample ID

Tested 09/03/24

Tester ENOR

Recorded 09/03/24

Lot # 4082315

Made 09/03/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager