



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: D4640 DICHLOROMETHANE HPLC (NOT STABILIZED)			
ACIDITY AS ACETIC ACID	0.0005 MAXIMUM WT%		<0.00002%
Sample ID			
Tested 09/04/24			
Tester TBUR			
Recorded 09/04/24			
APPEARANCE AND COLOR	CLEAR COLORLESS LIQUID APPEARANCE		CLEAR COLORLESS LIQUID FREE OF MAT
Sample ID			
Tested 09/04/24			
Tester TBUR			
Recorded 09/04/24			
ASSAY OF MATERIAL	MIN 99.9% %		>99.99%
Sample ID			
Tested 09/04/24			
Tester TBUR			
Recorded 09/04/24			
COLOR	MAX 10 APHA PT-CO		5 APHA
Sample ID			
Tested 09/04/24			
Tester TBUR			
Recorded 09/04/24			
RESIDUE OF EVAPORATION	MAX 4 PPM RESIDUEEVA		0.4 PPM
Sample ID			
Tested 09/04/24			
Tester TBUR			
Recorded 09/04/24			
FLUORESCENCE (AS QUININE BASE)	PASS TEST FLUORESCEN		PASSES TEST
Sample ID			
Tested 09/04/24			
Tester TBUR			
Recorded 09/04/24			
FREE HALOGENS	PASS TEST FREE HALOGEN		PASSES TEST
Sample ID			
Tested 09/04/24			
Tester TBUR			
Recorded 09/04/24			
GC/MS USE TEST	PASS TEST GC/MS USE		PASSES TEST
Sample ID			
Tested 09/04/24			
Tester TBUR			
Recorded 09/04/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

IDENTIFICATION	PASS TEST PASS	PASSES TEST
Sample ID		
Tested 09/04/24		
Tester TBUR		
Recorded 09/04/24		
REFRACTIVE INDEX	1.4215	1.4200 - 1.4230
	REFRACTIVE INDE	1.4212
Sample ID		
Tested 09/04/24		
Tester TBUR		
Recorded 09/04/24		
PERCENT WATER	MAX 0.02% % WATER	0.001%
Sample ID		
Tested 09/04/24		
Tester TBUR		
Recorded 09/04/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	07/30/2025
Sample ID		
Tested 09/04/24		
Tester TBUR		
Recorded 09/04/24		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested 09/04/24		
Tester TBUR		
Recorded 09/04/24		
MANUFACTURE DATE	MM/DD/YY MM/DD/YYYY	07/11/2024
Sample ID		
Tested 09/04/24		
Tester TBUR		
Recorded 09/04/24		
Lot # 424703		
Made 09/03/24		

Richard [Signature]

Quality Manager