



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: EQS510F			
MIXED METALS STANDARD 1.0 PPM CA, CR, CU, FE, MG, NA, NI, ZN			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 09/06/24			
Tester JTRA			
Recorded 09/06/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 09/06/24			
Tester JTRA			
Recorded 09/06/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P160,P157,P161,P390
Sample ID			
Tested 09/06/24			
Tester JTRA			
Recorded 09/06/24			
TRACEABLE TO N.I.S.T. STANDARD	YES		YES
Sample ID			
Tested 09/06/24			
Tester JTRA			
Recorded 09/06/24			
SHELF LIFE OF PRODUCT	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested 09/06/24			
Tester JTRA			
Recorded 09/06/24			
EXPIRATION DATE	MM/DD/YYYY		03/30/2025
Sample ID			
Tested 09/06/24			
Tester JTRA			
Recorded 09/06/24			
CALCULATED PPM CALCIUM	1.00 PPM	0.95 - 1.05	1.00
Sample ID			
Tested 09/06/24			
Tester JTRA			
Recorded 09/06/24			
CALCULATED PPM CHROMIUM	1.00 PPM	0.95 - 1.05	1.00
Sample ID			
Tested 09/06/24			
Tester JTRA			
Recorded 09/06/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

CALCULATED PPM COPPER (CU)	1.00	0.95 - 1.05	1.00
	PPM		
Sample ID			
Tested	09/06/24		
Tester	JTRA		
Recorded	09/06/24		
CALCULATED PPM IRON	1.00	0.95 - 1.05	1.00
	PPM		
Sample ID			
Tested	09/06/24		
Tester	JTRA		
Recorded	09/06/24		
CALCULATED PPM MAGNESIUM (MG)	1.00	0.95 - 1.05	1.00
	PPM		
Sample ID			
Tested	09/06/24		
Tester	JTRA		
Recorded	09/06/24		
Calculated PPM SODIUM	1.00	0.95 - 1.05	1.00
	PPM		
Sample ID			
Tested	09/06/24		
Tester	JTRA		
Recorded	09/06/24		
CALCULATED PPM NICKEL	1.00	0.95 - 1.05	1.00
	PPM		
Sample ID			
Tested	09/06/24		
Tester	JTRA		
Recorded	09/06/24		
CALCULATED PPM ZINC (ZN)	1.00	0.95 - 1.05	1.00
	PPM		
Sample ID			
Tested	09/06/24		
Tester	JTRA		
Recorded	09/06/24		
Lot #	4090505		
Made	09/06/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 09/06/24 at 10:49 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager