Page: 1

Date: 09/03/24 at 10:57 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	TOTAL CY	YANIDE BLIND-QA STANDA	RD (ALL SAME LOT)	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	09/03/24			
Tester	MMAT			
Recorded	09/03/24			
Actual ppm Cyanide		0.40	0.09 - 0.51	0.40
		PPM		
Sample ID				
Tested	09/03/24			
Tester	MMAT			
Recorded	09/03/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	09/03/24			
Tester	MMAT			
Recorded	09/03/24			
NSTRUMENTS USED D		INSTRUMENT		P291,GP,VF1-Z20
PREPARATION				, ,
Sample ID				
Tested	09/03/24			
Tester	MMAT			
Recorded	09/03/24			
EXPIRATION DATE	55, 66, ET	MM/DD/YYYY		09/17/2024
IM MANUE		MM/DD/YYYY		33/11/2024
Sample ID		W.W., 55, 1111		
Tested	09/03/24			
Tester	09/03/24 MMAT			
Recorded	09/03/24			
Lot # 40830				
Made 09/03	/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland In