



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 09/05/24			
Tester SNEL			
Recorded 09/05/24			
PH OF PRODUCT/SOLUTION	7.000 PH	6.000 - 12.500	10.149
Sample ID			
Tested 09/05/24			
Tester SNEL			
Recorded 09/05/24			
Actual Normality	0.1000 EQ/L	0.0995 - 0.1005	0.0997
Sample ID			
Tested 09/05/24			
Tester SNEL			
Recorded 09/05/24			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP10		YES
Sample ID			
Tested 09/05/24			
Tester SNEL			
Recorded 09/05/24			
QCSP35:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP35		YES
Sample ID			
Tested 09/05/24			
Tester SNEL			
Recorded 09/05/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 09/05/24			
Tester SNEL			
Recorded 09/05/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M113,M112,T56,M55
Sample ID			
Tested 09/05/24			
Tester SNEL			
Recorded 09/05/24			
Actual Density	1.01123 GM/ML	1.01079 - 1.01166	1.01127
Sample ID			
Tested 09/05/24			
Tester SNEL			
Recorded 09/05/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

DATE OF PREPARATION	MM/DD/YYYY	08/28/2024
Sample ID		
Tested	09/05/24	
Tester	SNEL	
Recorded	09/05/24	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	03/30/25
Sample ID		
Tested	09/05/24	
Tester	SNEL	
Recorded	09/05/24	
SHELF LIFE	6 MONTHS TIME	6 MONTHS
Sample ID		
Tested	09/05/24	
Tester	SNEL	
Recorded	09/05/24	
TRACEABLE TO NIST SRM 136	YES YES	YES
Sample ID		
Tested	09/05/24	
Tester	SNEL	
Recorded	09/05/24	
Lot #	4082641	
Made	09/05/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager