

Page: 1

Date: 09/05/24 at 6:01 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	: CY198F			
	AMMONIU	M NITROGEN STANDARD 5 P	PM AS NH3	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
CERTIFIED PPM NH3		5.0	4.0 - 6.0	5.0
		PPM		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
ALL EXITATION AND OC	LOIT	APPEARANCE		1 400
Sample ID		ALLEARANOL		
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24	INICTOLINACNIT		D000 VEF00 V44
INSTRUMENTS USED	DURING	INSTRUMENT		P396,VF500-X44
PREPARATION				
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
SHELF LIFE OF PRODUCT		6 MONTHS		6 MONTHS
		TIME		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
EXPIRATION DATE		MM/DD/YYYY		03/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	09/05/24			
Tester	AROS			
Recorded	09/05/24			
Lot # 4090514				
Made 09/05/24				
IVIAUE 397 0	·, - ·			



Page: 2

Date: 09/05/24 at 6:01 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**