

Made 09/09/24

Page: 1

Date: 09/09/24 at 11:17 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package	ed Product:	LY343F			
			STANDARD 100.0 PPM IN	2% HCL	
Test			Target/UOM	Range	Result
PREPARE	D TO FORMU	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	09/09/24			
	Tester	AROS			
	Recorded	09/09/24			
PPM NA			100.00	99.25 - 100.75	100.16
			PPM		
	Sample ID				
	Tested	09/09/24			
	Tester	AROS			
	Recorded	09/09/24			
APPEARANCE AND COLOR			CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID		7.11.12.11.01.11.01		
	Tested	09/09/24			
	Tester	AROS			
	Recorded	09/09/24			
INICTOLIMI			INSTRUMENT		M108,GC,VF1-Z25
INSTRUMENTS USED DURING PREPARATION		Oning	INSTITUMENT		W100,00, V1 1-223
PREPARA					
	Sample ID	00/00/04			
	Tested	09/09/24			
	Tester	AROS			
	Recorded	09/09/24	O MONTHO		O MONITUO
SHELF LII	FE OF PRODU	ICT	6 MONTHS TIME		6 MONTHS
	Sample ID				
	Tested	09/09/24			
	Tester	AROS			
	Recorded	09/09/24			
EXPIRATION	ON DATE		MM/DD/YYYY		03/30/2025
			MM/DD/YYYY		
	Sample ID				
	Tested	09/09/24			
	Tester	AROS			
	Recorded	09/09/24			
1.	ot # 40906				
LC)(# ⁴⁰³⁰⁰				



Page: 2

Date: 09/09/24 at 11:17 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager