Page: 1

Date: 09/09/24 at 2:35 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	HA-2076	50-32		
		ANADATE REAGENT (100mls	) FOR PHOSPHATES	
Test		Target/UOM	Range	Result
Absorbance of silica std vs blank		0.1110	0.1000 - 0.1220	0.1055
		ABSORBANCE		
Sample ID				
Tested	09/09/24			
Tester	TBUR			
Recorded	09/09/24			
Function test concentration blank		0.483-0.549 ABS		0.5071
		ABS		
Sample ID				
Tested	09/09/24			
Tester	TBUR			
Recorded	09/09/24			
Reagent blank		0.0185	0.0000 - 0.0370	0.0063
		ABSORBANCE(ABS)		
Sample ID				
Tested	09/09/24			
Tester	TBUR			
Recorded	09/09/24			
EXPIRATION DATE		MM/DD/YYYY		04/30/2029
		MM/DD/YYYY		
Sample ID				
Tested	09/09/24			
Tester	TBUR			
Recorded	09/09/24			
Lot # 42501	.3			
Made 09/06	/ 41			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager** 

hichard Stot