



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: OXY205F			
LACHAT CALIBRATION STANDARD LEVEL 5			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 09/10/24			
Tester JTRA			
Recorded 09/10/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 09/10/24			
Tester JTRA			
Recorded 09/10/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P58,M120,VF2-B14,M108
Sample ID			
Tested 09/10/24			
Tester JTRA			
Recorded 09/10/24			
PPM KIO3-NAOCL	20.0 PPM	19.9 - 20.1	20.0
Sample ID			
Tested 09/10/24			
Tester JTRA			
Recorded 09/10/24			
GPL SODIUM HYDROXIDE	100.00 GPL	99.95 - 100.05	100.00
Sample ID			
Tested 09/10/24			
Tester JTRA			
Recorded 09/10/24			
Actual Density	1.20471 GM/ML	1.20335 - 1.20606	1.20504
Sample ID			
Tested 09/10/24			
Tester JTRA			
Recorded 09/10/24			
CALCULATED GPL SODIUM CHLORIDE	180.00 GPL	179.98 - 180.02	180.00
Sample ID			
Tested 09/10/24			
Tester JTRA			
Recorded 09/10/24			
CALCULATED GPL SODIUM CHLORATE	0.5000 GPL	0.4995 - 0.5005	0.5000
Sample ID			
Tested 09/10/24			
Tester JTRA			
Recorded 09/10/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YYYY	09/30/2025
	MM/DD/YYYY	
Sample ID		
Tested	09/10/24	
Tester	JTRA	
Recorded	09/10/24	
Lot #	4090985	
Made	09/10/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager