

Page: 1

Date: 09/13/24 at 10:13 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: EV002F HYDOXYLAMINE HYDROCHLORIDE 0.04 MOLAR AQUEOUS SOLUTION				
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
	ple ID			
Teste	ed 09/13/24			
Teste	er JTRA			
Reco	orded 09/13/24			
ACTUAL MOLARI	TY	0.0400	0.0395 - 0.0405	0.0400
		MOLARITY (M/L)		
Samp	ple ID			
Teste				
Teste	er JTRA			
Reco				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sami	ple ID			
Teste				
Teste				
Reco				
INSTRUMENTS U		INSTRUMENT		M115,VF500-X30
PREPARATION	OLD DOMINO	IN OTH COME IN		W110, V1 000 700
	ple ID			
Teste				
Teste				
Reco	orded 09/13/24	0.99948	0.99928 - 0.99967	0.00053
Actual Density			0.99928 - 0.99967	0.99952
		GM/ML		
	ple ID			
Teste				
Teste				
Reco				
SHELF LIFE OF PRODUCT		6 MONTHS		6 MONTHS
		TIME		
	ple ID			
Teste				
Teste				
Reco				
EXPIRATION DAT	Έ	MM/DD/YYYY		03/30/2025
		MM/DD/YYYY		
Samp	ple ID			
Teste				
Teste				
Reco				
	091265			
	9/13/24			
Made ○	J/ IJ/ ZI			



Page: 2

Date: 09/13/24 at 10:13 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager