Page: 1

Date: 09/18/24 at 6:24 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	9597			
TOC STANDARD 2,500 PPM (KHP)				
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	09/13/24			
Tester	DTRA			
Recorded	09/13/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	09/13/24			
Tester	DTRA			
Recorded	09/13/24			
INSTRUMENTS USED DURING		INSTRUMENT		M81,GP,VF1-Z33
PREPARATION				
Sample ID				
Tested	09/13/24			
Tester	DTRA			
Recorded	09/13/24			
EXPIRATION DATE		MM/DD/YYYY		11/18/2024
		MM/DD/YYYY		
Sample ID				
Tested	09/13/24			
Tester	DTRA			
Recorded	09/13/24			
SHELF LIFE OF PRODUCT		TIME		2 MONTHS
Sample ID				
Tested	09/13/24			
Tester	DTRA			
Recorded	09/13/24			
Lot # 40912116				
Made 09/13				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stol