

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	FIS109F			
	Nitrate	Standard 2.0 ppm as NO3		
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	09/16/24			
Tester	JTRA			
Recorded	09/16/24			
CALCULATED PPM NI	TRATE (NO3)	2.00	1.95 - 2.05	2.00
		PPM W/V		
Sample ID				
Tested	09/16/24			
Tester	JTRA			
Recorded	09/16/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	09/16/24			
Tester	JTRA			
Recorded	09/16/24			
INSTRUMENTS USED DURING		INSTRUMENT		P137,VF500-X35
PREPARATION				
Sample ID				
Tested	09/16/24			
Tester	JTRA			
Recorded	09/16/24			
EXPIRATION DATE		MM/DD/YYYY		11/21/2024
		MM/DD/YYYY		
Sample ID				
Tested	09/16/24			
Tester	JTRA			
Recorded	09/16/24			
Lot # 4091	310			
Made 09/1				
maao				

hickard × -DI

Quality Manager