



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: DC503F			
SODIUM HYDROXIDE 2.77 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 09/16/24			
Tester DTRA			
Recorded 09/16/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 09/16/24			
Tester DTRA			
Recorded 09/16/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		09/14/2024
Sample ID	MM/DD/YYYY		
Tested 09/16/24			
Tester DTRA			
Recorded 09/16/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	17.0
Sample ID	DEG C		
Tested 09/16/24			
Tester DTRA			
Recorded 09/16/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		09/14/2024
Sample ID	MM/DD/YYYY		
Tested 09/16/24			
Tester DTRA			
Recorded 09/16/24			
AMBIENT TEST TEMPERATURE DEG C	20.0	15.0 - 30.0	17.0
Sample ID	DEG C		
Tested 09/16/24			
Tester DTRA			
Recorded 09/16/24			
Actual Normality	2.7700	2.7650 - 2.7750	2.7656
Sample ID	EQ/L		
Tested 09/16/24			
Tester DTRA			
Recorded 09/16/24			
QCSP7: STANDARDIZATION OF AQUEOUS BASES	YES		YES
Sample ID	METHOD		
Tested 09/16/24			
Tester DTRA			
Recorded 09/16/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 84 (KHP)	YES		YES
	YES		
Sample ID			
Tested	09/16/24		
Tester	DTRA		
Recorded	09/16/24		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	09/16/24		
Tester	DTRA		
Recorded	09/16/24		
Actual Density	1.10888	1.10843 - 1.10932	1.10844
	GM/ML		
Sample ID			
Tested	09/16/24		
Tester	DTRA		
Recorded	09/16/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,T62,M55
Sample ID			
Tested	09/16/24		
Tester	DTRA		
Recorded	09/16/24		
EXPIRATION DATE	MM/DD/YYYY		09/30/2026
	MM/DD/YYYY		
Sample ID			
Tested	09/16/24		
Tester	DTRA		
Recorded	09/16/24		
Lot #	4090606		
Made	09/16/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager