

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	1795			
-		NE STANDARD 1000 PPM AS	CL2 (Refrigerate)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES	~	YES
FILE		YES		
Sample ID				
Tested	09/16/24			
Tester	ENOR			
Recorded	09/16/24			
PPM CHLORINE (CL2)		1000.0	995.0 - 1005.0	1000.3
		PPM		
Sample ID				
Tested	09/16/24			
Tester	ENOR			
Recorded	09/16/24			
APPEARANCE AND CO		CLEAR PALE YELLOW		PASS
		APPEARANCE		
Sample ID				
Tested	09/16/24			
Tester	ENOR			
Recorded	09/16/24			
Actual Density	09/10/24	1.00100	0.99969 - 1.00230	1.00104
Actual Density		GM/ML	0.99909 - 1.00200	1.00104
Comunica ID		GW/WE		
Sample ID				
Tested	09/16/24			
Tester	ENOR			
Recorded	09/16/24			
		INSTRUMENT		M55,M120,M112,VF4-N32
PREPARATION				
Sample ID				
Tested	09/16/24			
Tester	ENOR			
Recorded	09/16/24			
EXPIRATION DATE		MM/DD/YY		11/21/24
		MM/DD/YYYY		
Sample ID				
Tested	09/16/24			
Tester	ENOR			
Recorded	09/16/24			
SHELF LIFE OF PRODU	ЈСТ	2 MONTHS		2 MONTHS
		TIME		
Sample ID				
Tested	09/16/24			
Tester	ENOR			
Recorded	09/16/24			
Lot # 40906	524			
Made 09/16)/24			



Page: 2 Date: 09/16/24 at 7:09 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager