



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8804			
SODIUM THIOSULFATE 0.0394 NORMAL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 09/19/24			
Tester ENOR			
Recorded 09/19/24			
<b>Actual Density</b>	1.00278	1.00261 - 1.00294	1.00287
	GM/ML		
Sample ID			
Tested 09/19/24			
Tester ENOR			
Recorded 09/19/24			
<b>QCSP10:STANDARDIZATION OF SODIUM THIOSUL</b>	YES		YES
	QCSP10		
Sample ID			
Tested 09/19/24			
Tester ENOR			
Recorded 09/19/24			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 09/19/24			
Tester ENOR			
Recorded 09/19/24			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M112,GC,T-293,BID-06969,PID -17
Sample ID			
Tested 09/19/24			
Tester ENOR			
Recorded 09/19/24			
<b>Normality Reported to 5 decimal</b>	0.03940	0.03932 - 0.03948	0.03939
	EQ/L		
Sample ID			
Tested 09/19/24			
Tester ENOR			
Recorded 09/19/24			
<b>PREPARED TO ASTM D 1510</b>	YES		YES
	YES		
Sample ID			
Tested 09/19/24			
Tester ENOR			
Recorded 09/19/24			
<b>SHELF LIFE OF PRODUCT</b>	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested 09/19/24			
Tester ENOR			
Recorded 09/19/24			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>EXPIRATION DATE</b>	MM/DD/YYYY	03/30/2025
	MM/DD/YYYY	
Sample ID		
Tested	09/19/24	
Tester	ENOR	
Recorded	09/19/24	
<b>Lot #</b>	40909108	
<b>Made</b>	09/19/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**