Page: 1

Date: 09/17/24 at 4:25 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	TES010F			
	NITROGEN 7.5 UG/0	(PPM) IN TOLUENE		
Test	Target/	UOM	Range	Result
PREPARED TO FORMULA			<b>J</b>	YES
FILE	YE			
Sample ID				
Tested (	09/17/24			
Tester	JTRA			
Recorded (	09/17/24			
<b>CERTIFIED PPM NITROG</b>	<b>EN:</b> 7.5		7.3 - 7.7	7.5
	PPI	I		
Sample ID				
	09/17/24			
	JTRA			
	09/17/24			
APPEARANCE AND COLO		3		PASS
	APPEAR			
Sample ID	7 = 7			
· ·	09/17/24			
	JTRA			
	09/17/24			
INSTRUMENTS USED DU		MENT		P46,VF1-Z6
PREPARATION	interne			1 40, 11 1 20
Sample ID				
•	09/17/24			
	J9/17/24 JTRA			
	)9/17/24			
EXPIRATION DATE	09/17/24 MM/DD/	/VVV		09/30/2025
EXPIRATION DATE	MM/DD/			09/30/2023
0 1 10	IVIIVI/DD/	111		
Sample ID	2011			
	09/17/24			
	JTRA			
	09/17/24			
Lot # 4091681				
<b>Made</b> 09/17/	24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I