09/18/24

Made

Page: 1

Date: 09/18/24 at 1:28 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Test		Target/UOM	Range	Result
	TO FORMULATION ON	YES	Range	YES
FILE	TO FORWIDEATION ON	YES		123
	Sample ID	120		
	Tested 09/18/24			
	Tester BFRA			
	Recorded 09/18/24	CLEAR LIQUID		DACC
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
	Sample ID			
	Tested 09/18/24			
	Tester BFRA			
	Recorded 09/18/24			
INSTRUMENTS USED DURING		INSTRUMENT		T77,T49,GC,M106
PREPARATI	ON			
	Sample ID			
	Tested 09/18/24			
	Tester BFRA			
	Recorded 09/18/24			
PH OF PRODUCT/SOLUTION		6.500	6.000 - 7.250	6.337
		PH		
	Sample ID			
	Tested 09/18/24			
	Tester BFRA			
	Recorded 09/18/24			
EXPIRATION		MM/DD/YY		12/30/24
	· PAIL	MM/DD/YYYY		12/00/24
	Sample ID	141141/ <i>DD</i> /1111		
	Tested 09/18/24			
	00/10/21			
	Tester BFRA Recorded 09/18/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St