



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9050			
SULFURIC ACID 0.2 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.2000 EQ/L	0.1995 - 0.2005	0.1997
Sample ID Tested 09/25/24 Tester LWIL Recorded 09/25/24			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID Tested 09/25/24 Tester LWIL Recorded 09/25/24			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		09/25/2024
PURPOSES	MM/DD/YYYY		
Sample ID Tested 09/25/24 Tester LWIL Recorded 09/25/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	17.9
Sample ID Tested 09/25/24 Tester LWIL Recorded 09/25/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		09/25/2024
Sample ID Tested 09/25/24 Tester LWIL Recorded 09/25/24			
AMBIENT TEST TEMPERATURE DEG C	20.0 DEG C	15.0 - 30.0	17.9
Sample ID Tested 09/25/24 Tester LWIL Recorded 09/25/24			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID Tested 09/25/24 Tester LWIL Recorded 09/25/24			
N.I.S.T. Traceable to SRM 723	Yes YES		YES
Sample ID Tested 09/25/24 Tester LWIL Recorded 09/25/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Density	1.00493 GM/ML	1.00479 - 1.00506	1.00490
Sample ID			
Tested	09/25/24		
Tester	LWIL		
Recorded	09/25/24		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	09/25/24		
Tester	LWIL		
Recorded	09/25/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	09/25/24		
Tester	LWIL		
Recorded	09/25/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		GC,T-293,M55
Sample ID			
Tested	09/25/24		
Tester	LWIL		
Recorded	09/25/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		09/30/26
Sample ID			
Tested	09/25/24		
Tester	LWIL		
Recorded	09/25/24		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		2 YEARS
Sample ID			
Tested	09/25/24		
Tester	LWIL		
Recorded	09/25/24		
Lot #	4091359		
Made	09/25/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 09/25/24 at 9:05 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager