



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: TEN033F			
SULFIDE STANDARD 2000 PPM			
Test	Target/UOM	Range	Result
PPM SULFIDE	2000 PPM	1980 - 2020	2011
Sample ID			
Tested	09/25/24		
Tester	JRIO		
Recorded	09/25/24		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	09/25/24		
Tester	JRIO		
Recorded	09/25/24		
QCSP56:STANDARDIZATION	YES		YES
SODIUM SULFIDE SO	QCSP56		
Sample ID			
Tested	09/25/24		
Tester	JRIO		
Recorded	09/25/24		
Actual Density	1.00908 GM/ML	1.00700 - 1.01116	1.01003
Sample ID			
Tested	09/25/24		
Tester	JRIO		
Recorded	09/25/24		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	09/25/24		
Tester	JRIO		
Recorded	09/25/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,VF1-Z36
Sample ID			
Tested	09/25/24		
Tester	JRIO		
Recorded	09/25/24		
SHELF LIFE OF PRODUCT	1 MONTH TIME		1 MONTH
Sample ID			
Tested	09/25/24		
Tester	JRIO		
Recorded	09/25/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		10/29/24
Sample ID			
Tested	09/25/24		
Tester	JRIO		
Recorded	09/25/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 09/25/24 at 11:33 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	4091019
Made	09/25/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager