

Made 09/27/24

Page: 1

Date: 09/27/24 at 3:32 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Customer PO #:				
Packaged Produc	t: 1866			
J		C ACID 1.0 MOLAR APHA FOR	RADIUM	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES	<u> </u>	YES
FILE		YES		
Sample II)			
Tested	09/27/24			
Tester	DTRA			
Recorded	09/27/24			
ACTUAL MOLARITY		1.0000	0.9950 - 1.0050	0.9974
		MOLARITY (M/L)		
Sample II)			
Tested	09/27/24			
Tester	DTRA			
Recorded				
N.I.S.T. Traceable to SRM 84		YES		YES
		YES		
Sample II				
Tested	09/27/24			
Tester	DTRA			
Recorded				P400
APPEARANCE AND COLOR		CLEAR TO PALE YELLOW		PASS
		LIQUID		
	_	APPEARANCE		
Sample II				
Tested	09/27/24			
Tester	DTRA			
Recorded	09/27/24	1.07654	1.07576 - 1.07732	1.07645
Actual Density		GM/ML	1.07576 - 1.07752	1.07645
Camania II	2	GIM/IVIL		
Sample II Tested				
Tester	09/27/24 DTRA			
Recorded				
NSTRUMENTS USED		INSTRUMENT		M112,VF5-L16,M55
PREPARATION	2011110	into into ment		
Sample II)			
Tested	09/27/24			
Tester	DTRA			
Recorded				
EXPIRATION DATE		MM/DD/YY		09/30/25
_		MM/DD/YYYY		
Sample II)			
Tested	09/27/24			
Tester	DTRA			
Recorded				



Page: 2

Date: 09/27/24 at 3:32 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager