

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	1794			
g		NE STANDARD 500 PPM AS (CL2	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	ILATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	10/02/24			
Tester	LWIL			
Recorded	10/02/24			
CHLORINE PPM		500 PPM	495 - 505	498
Sample ID				
Tested	10/02/24			
Tester	LWIL			
Recorded	10/02/24			
APPEARANCE AND COLOR		CLEAR PALE YELLOW APPEARANCE		PASS
Sample ID				
Tested	10/02/24			
Tester	LWIL			
Recorded	10/02/24			
Actual Density		1.00020 GM/ML	0.99955 - 1.00085	1.00020
Sample ID				
Tested	10/02/24			
Tester	LWIL			
Recorded	10/02/24			
INSTRUMENTS USED D		INSTRUMENT		M112,VF5-E5,P122,,BID00787
PREPARATION				,, ,,
Sample ID				
Tested	10/02/24			
Tester	LWIL			
Recorded	10/02/24			
TRACEABLE TO NIST S		YES		YES
		YES		
Sample ID				
Tested	10/02/24			
Tester	LWIL			
Recorded	10/02/24			
EXPIRATION DATE		MM/DD/YY MM/DD/YYYY		12/07/24
Sample ID				
Tested	10/02/24			
Tester	LWIL			
Recorded	10/02/24			
SHELF LIFE		2 MONTHS TIME		2 MONTHS
Sample ID				
Tested	10/02/24			
Tester	LWIL			
Recorded	10/02/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4093028
Made	10/02/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard St

Quality Manager