

Page: 1

Date: 10/08/24 at 11:30 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

Packaged Product				
	FLUORID	E STANDARD 20.0 PPM W/V	SOLUTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	ILATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	10/08/24			
Tester	JENC			
Recorded	10/08/24			
Actual ppm Fluoride		20.0	18.5 - 21.5	18.8
		PPM		
Sample ID				
Tested	10/08/24			
Tester	JENC			
Recorded	10/08/24			
QCSP45:MEASURING		YES		YES
WATER		QCSP45		
Sample ID				
Tested	10/08/24			
Tester	JENC			
Recorded	10/08/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID		, <u>_</u> ,		
Tested	10/08/24			
Tester	JENC			
Recorded	10/08/24			
INSTRUMENTS USED I		INSTRUMENT		GC,VF12-102,M118
PREPARATION		IIVO TITOMETYT		GG, VI 12 102, WI 10
Sample ID				
Tested	10/00/04			
Tested	10/08/24 JENC			
Recorded	10/08/24			
EXPIRATION DATE	10/00/24	MM/DD/YYYY		04/30/2025
EXPIRATION DATE		MM/DD/YYYY		04/30/2023
Sample ID				
Tested	10/08/24			
Tester	JENC			
Recorded	10/08/24			
Lot # 4100117				
Made 10/08/24				
mude ==7 °°	•			



Page: 2

Date: 10/08/24 at 11:30 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager