

Page: 1

Date: 10/07/24 at 8:42 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	EQS148F			
ŭ		C ACID 0.01 NORMAL		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	10/07/24			
Tester	ENOR			
Recorded	10/07/24			
Actual Normality		0.0100	0.0095 - 0.0105	0.0100
		EQ/L		
Sample ID				
Tested	10/07/24			
Tester	ENOR			
Recorded	10/07/24			
APPEARANCE AND CO	LOR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	10/07/24			
Tester	ENOR			
Recorded	10/07/24			
INSTRUMENTS USED D	URING	INSTRUMENT		P122,M113,BID-06969,PID-54
PREPARATION				
Sample ID				
Tested	10/07/24			
Tester	ENOR			
Recorded	10/07/24			
EXPIRATION DATE		MM/DD/YY		10/30/25
		MM/DD/YYYY		
Sample ID				
Tested	10/07/24			
Tester	ENOR			
Recorded	10/07/24			
SHELF LIFE OF PRODU	CT	SHELF LIFE		1 YEAR
		TIME		
Sample ID				
Tested	10/07/24			
Tester	ENOR			
Recorded	10/07/24			
Actual Density		0.99863	0.99843 - 0.99883	0.99861
		GM/ML		
Sample ID				
Tested	10/07/24			
Tester	ENOR			
Recorded	10/07/24			
Lot # 4100135				
Made 10/07	/24			



Page: 2

Date: 10/07/24 at 8:42 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager