



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> LY396F			
ORGANIC ACID CALIBRATION STOCK 1,000 PPM EA			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 10/08/24			
Tester MMAT			
Recorded 10/08/24			
<b>CALCULATED PPM PROPIONATE</b>	1000.00	995.00 - 1005.00	999.37
	PPM		
Sample ID			
Tested 10/08/24			
Tester MMAT			
Recorded 10/08/24			
<b>CALCULATED PPM ACETATE</b>	1000.00	995.00 - 1005.00	999.60
	PPM		
Sample ID			
Tested 10/08/24			
Tester MMAT			
Recorded 10/08/24			
<b>CALCULATED PPM FORMATE</b>	1000.00	995.00 - 1005.00	999.42
	PPM		
Sample ID			
Tested 10/08/24			
Tester MMAT			
Recorded 10/08/24			
<b>CALCULATED PPM BUTYRATE</b>	1000.00	995.00 - 1005.00	999.38
	PPM		
Sample ID			
Tested 10/08/24			
Tester MMAT			
Recorded 10/08/24			
<b>CALCULATED PPM CARBONATE</b>	1000.00	995.00 - 1005.00	999.90
	PPM		
Sample ID			
Tested 10/08/24			
Tester MMAT			
Recorded 10/08/24			
<b>CALCULATED PPM NITRATE</b>	1000.00	995.00 - 1005.00	999.83
	PPM		
Sample ID			
Tested 10/08/24			
Tester MMAT			
Recorded 10/08/24			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 10/08/24			
Tester MMAT			
Recorded 10/08/24			



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 10/08/24 at 4:22 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M108,VF500-X31
Sample ID Tested 10/08/24 Tester MMAT Recorded 10/08/24		
Lot # 4100865 Made 10/08/24		

**Quality Manager**