

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Proc	duct: SPX839F	n		
		Buffer Solution pH 10	.5	
Test		Target/UOM	Range	Result
PREPARED TO FO	ORMULATION ON	YES		YES
FILE		YES		
Samp	le ID			
Teste				
Teste				
Recor				
PH OF PRODUCT/SOLUTION		10.500	10.400 - 10.600	10.435
		PH		
Samp	le ID			
Teste				
Teste				
Recor	rded 10/18/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Samp	le ID			
Teste				
Teste				
Recor				
INSTRUMENTS USED DURING		INSTRUMENT		VF4-N30,EF4-YN,M108
PREPARATION				
Samp	le ID			
Teste				
Teste				
Recor				
EXPIRATION DATE		MM/DD/YYYY		02/28/2025
		MM/DD/YYYY		
Samp	le ID			
Teste				
Teste				
Recor				
Lot # 4	101529			
	0/18/24			
maac				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA