

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product: 6966	5		
	DUCTIVITY STANDARD 84.0 UMH	IOS @25 C	
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION O			YES
FILE	YES		
Sample ID			
Tested 10/18/24			
Tester BPHI			
Recorded 10/18/24			
Certified Conductivity @25 degree	es 84.0	83.9 - 84.1	84.0
C	UMHOS		
Sample ID			
Tested 10/18/24			
Tester BPHI			
Recorded 10/18/24	YES		YES
QCSP30:CONDUCTIVITY			YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 10/18/24			
Tester BPHI			
Recorded 10/18/24			
TRACEABLE TO N.I.S.T. SRM 319	1 YES		YES
	YES		
Sample ID			
Tested 10/18/24			
Tester BPHI			
Recorded 10/18/24			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID	/		
Tested 10/18/24			
Tester BPHI			
Recorded 10/18/24			
INSTRUMENTS USED DURING	INSTRUMENT		M92,T-234
	INSTRUMENT		10192,1-234
PREPARATION			
Sample ID			
Tested 10/18/24			
Tester BPHI			
Recorded 10/18/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		10/30/25
Sample ID			
Tested 10/18/24			
Tester BPHI			
Recorded 10/18/24			
SHELF LIFE OF PRODUCT	1 YEAR		1 YEAR
	TIME		
Sample ID			
Tested 10/18/24			
Tester BPHI			
Recorded 10/18/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4101853
Made	10/18/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

thand St

Quality Manager