



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4090			
HYDROCHLORIC ACID 0.01 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 10/17/24			
Tester JRIO			
Recorded 10/17/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 10/17/24			
Tester JRIO			
Recorded 10/17/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		10/17/2024
Sample ID			
Tested 10/17/24			
Tester JRIO			
Recorded 10/17/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	20.0
Sample ID			
Tested 10/17/24			
Tester JRIO			
Recorded 10/17/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		10/17/2024
Sample ID			
Tested 10/17/24			
Tester JRIO			
Recorded 10/17/24			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	20.0
Sample ID			
Tested 10/17/24			
Tester JRIO			
Recorded 10/17/24			
Actual Normality	0.0100	0.0095 - 0.0105	0.0098
Sample ID			
Tested 10/17/24			
Tester JRIO			
Recorded 10/17/24			
+ Measurement Uncertainty	0.0001	= 0.0001	0.0001
Sample ID			
Tested 10/17/24			
Tester JRIO			
Recorded 10/17/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

**QCSP5: STANDARDIZATION OF
 AQUEOUS ACIDS**

YES
 METHOD YES

Sample ID
 Tested 10/17/24
 Tester JRIO
 Recorded 10/17/24

N.I.S.T. Traceable to SRM 723 THAM

YES
 YES YES

Sample ID
 Tested 10/17/24
 Tester JRIO
 Recorded 10/17/24

APPEARANCE AND COLOR

CLEAR LIQUID
 APPEARANCE PASS

Sample ID
 Tested 10/17/24
 Tester JRIO
 Recorded 10/17/24

Actual Density

0.99842
 G/ML 0.99832 - 0.99852 0.99838

Sample ID
 Tested 10/17/24
 Tester JRIO
 Recorded 10/17/24

**INSTRUMENTS USED DURING
 PREPARATION**

INSTRUMENT M115,VF12-101

Sample ID
 Tested 10/17/24
 Tester JRIO
 Recorded 10/17/24

EXPIRATION DATE

MM/DD/YY
 MM/DD/YYYY 10/30/25

Sample ID
 Tested 10/17/24
 Tester JRIO
 Recorded 10/17/24

SHELF LIFE OF PRODUCT

1 YEAR
 TIME 1 YEAR

Sample ID
 Tested 10/17/24
 Tester JRIO
 Recorded 10/17/24

Lot # 4101056

Made 10/17/24



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 10/17/24 at 2:17 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager